

# ASSAY CERTIFICATE

**ANTECH LABORATORY**  
**P. O. BOX 150, KWEKWE**  
**6 KM PEG MVUMA ROAD**



Tel: +263 55 22269/24172/22682

Cell: +263 772 864 065

Email: contact@antechlaboratories.com

Edson.Luwobo@antechlaboratories.com

ISO/IEC 17025: 2017

A SADCAS Accredited Testing Laboratory: TEST-5 0030

**Mpofini Mine**  
**Tshotshe Area**  
**Silobela**  
**0727 719 093**

Date Delivered : 02.03.2026  
 Date Logged : 02.03.2026  
 Your Ref : 02.03.2026  
 Our Ref : MOP 7215  
 Sample Description : Dumps  
 Legend : parts per million [ppm]  
 ND : Not Detectable  
 Analysis : Bottle Roll with AA Finish [Detection Limit 0.01 MT5]  
 Elements : Au  
 Number of Samples : 01  
 Date Analysed : 03.03.2026  
 Date Reported : 03.03.2026

Analyst : M.Chipindu  
 Instrument : Agilent 240 FS AA

### Recommended Limits for Laboratory Standard Samples:

Ref: 1 Standard 3 : 0.26 – 0.34  
 Ref: 2 ANT 25/2 Standard : 1.30 – 1.86  
 Ref: 3 Blank Standard : <0.01

Lab ID	Customer Sample ID	Au ppm
ANT 25/2 Standard		1.44
ppm		
X7120 MPOFINI MINE		0.09
ppm		
Standard 3		0.29
ppm		
Blank Standard		<0.01
ppm		

<b>ANTECH LABORATORY</b>	
P.O. BOX 150, KWEKWE	
TELEPHONE: 22259/24172	
MANAGER: .....	
SIGNATURE: .....	
DATE: .....	5/3/2026

Reviewed by		Authorized by	
Technical Signatory	Date	Technical Signatory	Date
M. Kanjanda	5/3/2026	M. Kanjanda	5/3/2026

[1] Only hardcopy Assay Certificate authorised by the Technical Signatory are to be considered valid.

[2] Report only to be copied in full.

[3] The Laboratory is not responsible for the customer's sampling errors in the field.

[4] Results only relate to sample received & tested.

[5] Disclaimer: Results marked [AMNA] are excluded from ISO Accreditation.

[6] Course rejects will be stored for one month while Pulp rejects will be stored for six months.